

ENQUIRY

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Description	Qty	UOM
1	<p>Scan vac Speed vacuum concentration system (SS and teflon coated)with lyophilizer.</p> <p>Specifications:</p> <ol style="list-style-type: none"> 1. The system should consist of vacuum centrifuge and freezer dryer. 2. The system should have rotors for 1.5/2ml,0.5ml eppendorf tubes. 3. The system should rpm range of 0-2000rpm. 4. The system should have brushless dc motor drive. 5. The chamber volume of freezer dryer should be 3 liters. 6. The system cold trap should be able to attain the temperature of-100oC. 7. The system should have minimum 6 ports for attaching flasks. 8. The system should have programs to store/edit different protocols. 9. The system should be able to attain vacuum level of 0.001mbar. 10. The system should have option to connect for N2 to purge N2 while release the vacuum. 11. The system should have LCD display for displaying temperature, vacuum level,RPM and time. 12. The system should have timer range of 0-999 minutes 13. The system should have keypad control panel for changing the set parameter as well as to clear the errors. 14. The control panel should have the interlocking between temperature and vacuum and also it should give the alarms for power failure, water in condenser and service due date etc. 15. The system should be supplied with heavy duty vacuum pump for 24*7 operation with spare oil required for the pump. 16. The system should have RS-232 or USB port to transmit data to a user-supplied computer. 17. The system should have drain hose with valve in order to remove the water from chamber. 18. The system should be supplied with all the required accessories to function. 19. The system must have 230 volt reverse IEC plug at the back. 20. The system must have fuses for protection fuses of vacuum pump and main unit. 21. The warranty for the system should be 3 years. 22. The system should be CE/ISO certified. 	1.00	NO

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 162826. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 20/12/2017

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences



SAMPATTU K S