



National Center for Biological Sciences
Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6346

Email: nalinir@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/CA18-174/181920

January 16, 2019

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Code	Description	Qty	UOM
1		<p>TECHNICAL SPECIFICATION for Cooled incubator</p> <p>The system should have the capacity of 120 liters \pm 5% The system should be microprocessor controlled Incubator It should have LCD Display with easy icons and user-friendly control system. Should have a Temperature Control Range: -10°C to $+60^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ The system Should have Temperature Uniformity of $\pm 0.5^{\circ}\text{C}$ at 35°C The system Should have Thermistor Temperature Sensor The system should have the Programmable operation function with microprocessor control for lighting on/off control The system should have the Air circulation by a fan to maintain temperature uniformity The system should have the Over-temperature protection Device Should have 40mm Thickness of insulation with PUF Insulation Material The internal material of construction should be SS 304. The should have an option to mute the audible alarms The system should have the Corrosion resistance stainless steel interior and epoxy-powder coated exterior. Should have minimum three adjustable shelves Should have One Access Port for easy communication The system should have Internal Fluorescent Lamp with Timer Electronic and mechanical components should be enclosed, protected from accidental spills. It should have automatic restart after power interruption with nonvolatile memory The system should operate at 230V, 50 Hz AC power. The system should have Noise Level less than 40 Db The system should be supplied with all the accessories required to function. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required). The system should be supplied with 3 years warranty. The System should be CE/ISO Approved</p>	1.00	NO

NOTE :

Kindly provide Authorised Distributor certificate and Proprietary certificate

The tender can be quoted in all tradable currencies such as USD, EURO, GBP, etc

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for Ref No: 181920. The bids should be addressed and to be mailed/couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE? with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 30/01/2019 BY 5.30PM.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.

14. TIFR is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.

15. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

16. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services

Yours faithfully

For and on behalf of

National Centre for Biological Sciences



Head Purchase