



National Center for Biological Sciences
Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6347

Email: shabeentaj@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/IC17-167/163126

December 20, 2017

Dear Sirs,

Please let us have your Quotation for the following :

S.No.	Item Code	Item Description	Qty	UOM
1	FCQFR000059	-20°C Freezer as per specification attached- Annexure-A	1.00	NO
2	FCQFR000060	4°C Freezer as per specification attached- Annexure-B	1.00	NO

Please also indicate specifically the following:

1. The bids / quotation addressed to "THE HEAD – PURCHASE" with Tender / Enquiry Ref No. and Item description should reach us before submission date and time. The bids/quotation are liable to be rejected, if not addressed to "THE HEAD – PURCHASE".

2. EXIM CODE - 8 digit Customs Tariff Number – HSN (Harmonised System Nomenclature)

3. Country of origin

4. Place of shipment

5. Name, address, telephone number and fax numbers of your bankers

6. Your account number and swift number with the bank (required for payment purposes if the order materialises on your firm)

7. Please indicate freight and insurance charges separately.

8. Please indicate the Warranty Period and Delivery Period

9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by 08/01/2018

Important Note:

10. Shipping by AIRFREIGHT ONLY – CARGO MODE
11. Port of Entry for Customs clearance is – BANGALORE
12. Please use International Freights which arrive directly into Bangalore. Ex. Lufthansa Airlines Cargo, Singapore Airlines Cargo
13. NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX, TNT, UPS, DHL ETC. SHIPMENTS SENT THROUGH FEDEX, TNT, UPS, DHL, ETC. WILL NOT BE ACCEPTED BY NCBS.
14. NCBS is exempted in paying customs duty (except advalorem duty of 5% + 2% cess and 1% Cus Sec & High Edu.CESS on CIF value)

PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK (CASH AGAINST DOCUMENTS). DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION. BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT. ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully
For and on behalf of
National Centre for Biological Sciences



SHABEEN TAJ

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. HEMANT SOPAN PEDDAWAD (NCB-12P4510)
Your Indent Req. No.TEM2700005 Dated:18/12/2017 refers.

NCB/163126/17-18 (T)

Specification for -20 C Freezer- Annexure-A

1. System should be vertical standing and should have working volume of 600 liters +/- 5%
2. The system should be a research grade and suitable for continuous operation.
3. The system should have digital display and Micro controller based user-friendly control system.
4. System should be equipped with precise temperature controller with accuracy $\pm 1^{\circ}\text{C}$ & it should maintain uniform temperature in freezer.
5. The freezer/Fridge should have single door.
6. The system should have manual defrost.
7. The preferred color should be white.
8. The temperature should be stable at -20 degrees
9. The system should have suitable display for temperature.
10. The system should have minimum 12 shelves for storage of samples.
11. The freezer should have audio- visual alarms for low/ high temperature, door open and power failure. there must be an option to mute the audible alarms.
12. It should be supplied with stabilizer and all other accessories required for proper functioning of freezer.
13. The system should have efficiency and low power consumption.
14. The system should have high grade thermal insulation for freezer walls.
15. Power supply should be 230V, 50HZ
16. The system should have Caster moving wheels with stopper for easy movement
17. The system should have minimum 3-year warranty
18. The system should be CE/ISO certified.
19. The Company must quote the freezer with all the suitable documents to justify the specification of product.

Specification of 4°C Refrigerator- Annexure-B

1. System should be vertical standing and should have working volume of 500 Ltrs ($\pm 5\%$)
2. The system should be a research grade and suitable for continuous operation.
3. System should be equipped with precise temperature controller with accuracy $\pm 1^\circ\text{C}$ & it should maintain uniform temperature in freezer.
4. The system should have manual defrost.
5. The system should have double sliding door.
6. The system should have 6 shelves.
7. The system should have audio- visual alarms for low/ high temperature, door open and power failure. Preferably there should be an option to mute the audible alarms.
8. It should be supplied with stabilizer and all other accessories required for proper functioning of refrigerator.
9. The system should have Caster moving wheels with stopper for easy movement
10. The system should have minimum 3-year warranty
11. The system should be CE/ISO certified.
12. Power supply should be 230V, 50HZ
13. The system should have efficiency and low power consumption.
14. The system should be high power efficient, have high grade thermal Insulation.
15. The system should have suitable display for temperature.
16. The system should be certified for research grade as per norms of IS & CE or equivalent.
17. The Company must quote the refrigerator with all the suitable documents to justify the specification of product.