



National Center for Biological Sciences
Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6347

Email: shabeentaj@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/CA17-164/162875

November 28, 2017

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Code	Description	Qty	UOM
1	LCQFR000038	4 Degree Refrigerator (Asper attached specification)	1.00	NO

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 162875. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 18/12/2017.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences


SHABEEN TAJ

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. AWADHESH PANDIT (NCB-12P4510)
Your Indent Req. No. APA1370034 Dated: 22/11/2017 refers.

Specification for 4° Refrigerator

1. EffectiveCapacity : 700ltrs ± 20 ltrs
2. Exterior : Galvanized steel with baked-on finish
3. Interior : Galvanized steel with baked-on finish
4. Door : 1 x double pane glass doors, self-closing
5. Insulation : HCFC-free rigid foamed-in-place polyurethane
6. Shelves: 5 x coated steel drawers, Handles with card holder.
7. Access ports : 3 locations, 30mm diameter (both sides, cabinet top)
8. Lock : 1 x cylinder lock
9. Casters : 4pcs
10. Cooling Method : Forced air circulation with single fan
11. Defrosting Method : Forced type (cycle defrost system), fully automatically hot pipe for automatically evaporation for drain water
12. Heater : 153W
13. Temperature control: Microprocessor controlled on/off type.
14. Temperature sensor : Thermistor
15. Temperature range : 2°C to 14°C (ambient temperature - 5°C to 0°C no-load) 2°C to 23°C
16. (ambient temperature - 0°C to 35°C no-load) (1°C step)
17. Safety features: High/low temperature protection circuit, remote alarm contact, temperature lock, part replacement notification, self-diagnostics, memory backup (non-volatile memory).
18. High Temperature alarm : Flashing LED and buzzer after 15 minutes
19. Low Temperature alarm : Flashing LED and buzzer after 15 minutes
20. Door ajar alarm : Flashing lamp, buzzer after 0 to 15 minutes delay
21. Interior lamp : 20W x 1, fluorescent lamp