



National Center for Biological Sciences
Tata Institute of Fundamental Research
GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6347
Email: shabeentaj@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/CA17-249/163303

January 06, 2018

Dear Sirs

Please let us have your Quotation for the following:

| S.No. | Item Code | Description | Qty | UOM |
|-------|-------------|--|------|-----|
| 1 | LCQBE000226 | 8 channel Remote alarm system for -80,-20 and CO2 Incubators (As per the specification attached) | 2.00 | NOS |
| 2 | LCQBE000227 | 4 channel Remote alarm system for -80,-20 and CO2 Incubators (As per the specification attached) | 1.00 | NO |

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 163303. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 25/01/2018.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences



SHABEEN TAJ

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. JAYAPRAKASH R. (NCB-12P4510)
Your Indent Req. No. JRA1200120 Dated: 22/12/2017 refers.

Remarks: Remote alarm system for -80, -20 and CO2 Incubators



national centre for biological sciences
tata institute of fundamental research

Ref: NCB/163303/17-18 (T)

Specification of Remote alarm system

1. System should have 8 input channels
2. System should have telephone based communication.
3. System should have minimum eight custom voice call alarm notification system
4. The system should have inbuilt microphone to record the customized voice messages.
5. System should be able to connect with -80 Freezers, -20 freezers and CO2 Incubators to generate the alarm in case of any failure in the Equipment (power, Temperature and CO2 %)
6. The system should be wall or desktop mounted.
7. System should be a fully programmable monitoring system.
8. The system should have option to add/delete the programming.
9. System should have inbuilt battery backup
10. Power 230V 50 HZ
11. System should have 3 years warranty
12. System should have CE/ISO Certified
13. System should be supplied with all the accessories required to function