

ENQUIRY

Dear Sirs,

Please let us have your Quotation for the following:

S.No.	Item Code	Item Description	Qty	UOM
1	FCQMI000142	Refrigerated Centrifuge <u>Specification of Table top Refrigerated Centrifuge</u> 1. The system should have a maximum speed of 14000 RPM. 2. System should be suitable for high capacity rotors upto 750ml and swing bucket rotors/plate rotors 3. The system should have LCD Display with Microprocessor based control System. 4. The system should have Direct, Brush less induction low profile motor drive system 5. The system should have Imbalance Detection system. 6. The system should have Soft-touch one-finger lid closure for easy operation 7. The system should have temperature ranging from -10° C to 40° C and quick precooling 8. The system should have Pre-cooling system with Direct Button. 9. The system should have Timer Range from 35 secs to 9:50 hrs, and continuous run function. 10. The system should have Automatic lid opening at the end of the run to prevent the sample and allows easy access to samples 11. The system should should have Power failure and Temperature Alarm 12. The system should have the Built-in condensation chamber to avoid water accumulation in the rotor chamber 13. The system Should have soft brake to protect sensitive sample. 14. The system Should have option for auto power shut off after 8 hr of non- use to extend the compressor life. 15. The system should be power efficient and low power consumption. 16. The system should be supplied with all the accessories required to function. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required). 17. The system should be supplied with 3 years warranty. 18. The system should be CE/ISO Certified 19. The offer should include the cost of rotors as optional items	1.00	NO
2	FCQMI000140	Spinwin Centrifuge Variable Speed 6 places, 0.5/2ml tubes Temperature Range - Room temperature Latch lid The system should be CE/ISO Certified	2.00	NOS

3.	FCQMI000143	Refrigerated centrifuge with 1.5ml/2ml rotor	2.00	NOS
<p>Specification of Table top Refrigerated Centrifuge</p> <ol style="list-style-type: none"> 1. The system should have a maximum speed of 14000 RPM. 2. Rotor capacity should 24 x 1.5ml/2 ml tubes 3. The system should have LCD Display with Microprocessor based control System. 4. The system should have Direct, Brush less induction low profile motor drive system 5. The system should have Imbalance Detection system. 6. The system should have Soft-touch one-finger lid closure for easy operation 7. The system should have temperature ranging from -10° C to 40° C quick precooling 8. The system should have Pre-cooling system with Direct Button. 9. The system should have Timer Range from 35 secs to 9:50 hrs, and continuous run function. 10. The system should have Automatic lid opening at the end of the run to prevent the sample and allows easy access to samples 11. The system should have Power failure and Temperature Alarm 12. The system should have the Built-in condensation chamber to avoid accumulation in the rotor chamber 13. The system should have soft brake to protect sensitive sample. 14. The system should have option for auto power shut off after 8 hr or more use to extend the compressor life. 15. The system should be power efficient and low power consumption 16. The system should be supplied with all the accessories required to function. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochure should be highlighted wherever required). 17. The system should be supplied with 3 years warranty. 18. The system should be CE/ISO Certified 				

NOTE : **TIFR is a public funded research institute and is entitled to concessional rate of IGST @ 5% for items supplied for research purpose vide notification no. 47/2017 dated 14th Nov, 2017 . The offer should be submitted after fully considering the above notification. **

Remarks: The Tender Can be quoted in all Tradable currencies such as USD,GBP, EURO, Yen, Etc.,

Please also indicate specifically the following:

1. The bids / quotation addressed to “THE HEAD – PURCHASE” with Tender / Enquiry Ref No. and Item description should reach us before submission date and time. The bids/quotation are liable to be rejected, if not addressed to “THE HEAD – PURCHASE”.
2. **EXIM CODE - 8 digit Customs Tariff Number – HSN (Harmonized System Nomenclature)**
3. Country of origin
4. Place of shipment
5. Name, address, telephone number and fax numbers of your bankers
6. Your account number and swift number with the bank (required for payment purposes if the order materialises on your firm)
7. **Please indicate freight and insurance charges separately.**
8. Please indicate the Warranty Period and Delivery Period
9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by 02/01/2018

National Center for Biological Sciences

Tata Institute of Fundamental Research
GKVK, Bellary Road, Bangalore 560 065. India
Email: nalinir@ncbs.res.in, www.ncbs.res.in
Tel. No: 080 2366, 6346

12. Please use International Freights which arrive directly into Bangalore. Ex. Lufthansa Airlines Cargo, Singapore Airlines Cargo
13. NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX, TNT, UPS, DHL ETC. SHIPMENTS SENT THROUGH FEDEX, TNT, UPS, DHL, ETC. WILL NOT BE ACCEPTED BY NCBS.
14. NCBS is exempted in paying customs duty (except advalorem duty of 5% + 2% cess and 1% Cus Sec & High Edu.CESS on CIF value)
15. PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK (CASH AGAINST DOCUMENTS). DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION. BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT. ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully
For and on behalf of
National Centre for Biological Sciences



NALINI R

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. P.P.RANJITH (NCB-12P4510)
Your Indent Req. No.PPR0240282 Dated: 07/12/2017 refers.

