

ENQUIRY

December 22, 2017

Ref: NCB/IC17-158/163063

Dear Sirs,

Please let us have your Quotation for the following :

S.No.	Item Code	Item Description	Qty	UOM
1	FCQMC000067	<p>Imaging system compatible with</p> <ol style="list-style-type: none"> 1. motorized MBF stage and controller. 2. MBF Z encoder. 3. Ludle joy stick <p>in the item details section.</p> <p>General Specifications</p> <p>Frame 12V100W Koehler illumination system, Coarse & Fine focusing knobs, built in LED light for light intensity display, preset button switch, with gradation filter and built-in filter holders for three filters (LBD, ND6 and ND25), Motorized reflected light fluorescence illuminator F.N 22 equipped with motorized shutter, field stop, aperture stop and motorized 8-position mirror unit turret, including ND filter slider AW2176, UV protective shield JA5311, shield plate JA5290, allen wrench AD6162, indication plate JA5820, and shield cap AC0792 (4 pcs) Corresponded to RoHS and WEEE regulations</p> <p>Mirror Unit for fluorescence microscopy, wide band UV excitation, exciter filter BP340-390, dichroic beamsplitter DM410, barrier filter BA420-IF, including plastic case AR1924</p> <p>Transmitted light system with 100W halogen light source</p> <p>Coded function enabled 6 position objective holder with control unit</p> <p>Observation tube with photoport and three position prism for selection of light path for 100% observation, 20% - 80% Camera Port - observation, 100% Camera Port. 10X Magnification eyepieces of F.N 22</p> <p>Dual port adapter for 2 Cameras, Swing out type light condenser. Optics with 4 (N.A 0.13), 10 (N.A 0.3), 20 (N.A 0.5), 40 (N.A0.75) and air immersion 100(N.A 1.3) times magnification with the highest magnification should be a Achromat type.</p>	1.00	NO

S.No.	Item Code	Item Description	Qty	UOM
		<p>Optics based observation and Imaging Frame for the existing MBF Neuronal imaging and analysis system and should be compatible with existing Motorized XY stage and other hardwares.</p> <p>Should be compatible with stage controller, Z motor and camera and support the existing software supplied by MBF bioscience.</p>		

Please also indicate specifically the following:

1. The bids / quotation addressed to "THE HEAD – PURCHASE" with Tender / Enquiry Ref No. and Item description should reach us before submission date and time. The bids/quotation are liable to be rejected, if not addressed to "THE HEAD – PURCHASE".
2. EXIM CODE - 8 digit Customs Tariff Number – HSN (Harmonised System Nomenclature)
3. Country of origin
4. Place of shipment
5. Name, address, telephone number and fax numbers of your bankers
6. Your account number and swift number with the bank (required for payment purposes if the order materialises on your firm)
7. Please indicate freight and insurance charges separately.
8. Please indicate the Warranty Period and Delivery Period
9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by 05/01/2018

Important Note:

10. Shipping by AIRFREIGHT ONLY – CARGO MODE
11. Port of Entry for Customs clearance is – BANGALORE
12. Please use International Freights which arrive directly into Bangalore. Ex. Lufthansa Airlines Cargo, Singapore Airlines Cargo
13. NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX, TNT, UPS, DHL ETC. SHIPMENTS SENT THROUGH FEDEX, TNT, UPS, DHL, ETC. WILL NOT BE ACCEPTED BY NCBS.
14. NCBS is exempted in paying customs duty (except ad valorem duty of 5% + 2% cess and 1% Cus Sec & High Ed. CESS on CIF value)