



National Center for Biological Sciences
Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6343

Email: nirmala@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/IC17-171/163155

December 22, 2017

Dear Sirs,

Please let us have your Quotation for the following :

S.No.	Item Code	Item Description	Qty	UOM
1	FCQBE000211	Table top Centrifuge with rotors.1. The system should have a maximum speed of 14000 RPM. 2. The system should have LCD Display with Microprocessor based control System. 3. The system should have Direct, Brushless induction, low profile motor drive system 4. The system should have Imbalance Detection system. 5. The system should have Soft-touch one-finger lid closure for easy operation 6. The system should have temperature control ranging from -10° C to 40° C and quick precooling 7. The system should have Pre-cooling system with Direct Button. 8. The system should have Timer Range from 35secs to 9:50hrs, and continuous run function. 9. The system should have Automatic lid opening at the end of the run to prevent the sample and allow easy access to samples 10. The system should have Power failure and Temperature Alarm 11. The system should have the Built-in condensation chamber to avoid water accumulation in the rotor chamber 12. The system should have soft brake to protect sensitive samples. 13. The system should have option for auto power shut off after 8 hr of non- use to extend the compressor life. 14. The system should be power efficient and low power consumption. 15. The system should be supplied with all the accessories required	2.00	NOS

S.No.	Item Code	Item Description	Qty	UOM
		to function. Compliance to each of the above points should be separately indicated and evidence presented for each of them (Product brochures should be highlighted wherever required). 16. The supplier shall be responsible to pay the amount for any damage to the sample due to breakdown during warranty period. 17. The system should be CE/ISO Certified 18. System should be suitable for high capacity rotors upto 750ml and swinging bucket rotors/plate rotors.		

Please also indicate specifically the following:

1. The bids / quotation addressed to "THE HEAD - PURCHASE" with Tender / Enquiry Ref No. and Item description should reach us before submission date and time. The bids/quotation are liable to be rejected, if not addressed to "THE HEAD - PURCHASE".

2. EXIM CODE - 8 digit Customs Tariff Number - HSN (Harmonised System Nomenclature)

3. Country of origin

4. Place of shipment

5. Name, address, telephone number and fax numbers of your bankers

6. Your account number and swift number with the bank (required for payment purposes if the order materialises on your firm)

7. Please indicate freight and insurance charges separately.

8. Please indicate the Warranty Period and Delivery Period

9. Kindly expedite sending your Quote/Proforma Invoice/Invoice by courier so as to reach us latest by 05/01/2018

Important Note:

10. Shipping by AIRFREIGHT ONLY - CARGO MODE

11. Port of Entry for Customs clearance is - BANGALORE

12. Please use International Freights which arrive directly into Bangalore. Ex. Lufthansa Airlines Cargo, Singapore Airlines Cargo

13. NO SHIPMENT SHOULD BE MADE THROUGH COURIER COMPANIES LIKE FEDEX, TNT, UPS, DHL, ETC. SHIPMENTS SENT THROUGH FEDEX, TNT, UPS, DHL, ETC. WILL NOT BE ACCEPTED BY NCBS.

14. NCBS is exempted in paying customs duty (except advalorem duty of 5% + 2% cess and 1% Cus Sec & High Edu. CESS on CIF value)

PLEASE NOTE THAT PAYMENT WILL BE MADE THROUGH SIGHT DRAFT THROUGH BANK (CASH AGAINST DOCUMENTS). DOCUMENTS WILL BE RELEASED FROM THE BANK IMMEDIATELY ON RECEIPT OF THE INTIMATION. BANK CHARGES INSIDE INDIA SHALL BE TO NCBS ACCOUNT AND OUTSIDE INDIA SHALL BE TO YOUR ACCOUNT. ALSO DO NOT INCLUDE CHARGES TOWARDS CUSTOMS DUTY IN YOUR PRICES SINCE NCBS WILL BEAR THIS EXPENDITURE.

Yours faithfully

For and on behalf of

National Centre for Biological Sciences

NIRMALA K.S.

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. NAVEEN KUMAR (NCB-12P4510)
Your Indent Req. No. TEM2120005 Dated: 14/12/2017 refers.